Authors have nothing to disclose with regard to commercial support.

EARLY DEGENERATION OF THE ST JUDE MEDICAL TRIFECTA BIOPROSTHETIC AORTIC VALVE: A PROBLEM OF

THE LEAFLETS OR OF THE STENT?

To the Editor:

The development of new aortic valvular bioprosthetic devices is a very important issue in light of the increasing demand for these models for younger patients who want to avoid oral anticoagulation and maintain a better quality of life. Moreover, the prospect of a subsequent catheter-based valve-in-valve therapy plays an even more important role in this new trend in decision making.

If we are to implant these devices in younger and younger patients, the aspect of long-term durability becomes central in the development of the new models. With regard to this issue, we read with interest the article by Goldman and colleagues.¹ They reported on a prospective, multicenter trial in which 710 patients operated on for aortic valve disease received the novel Trifecta (St Jude Medical, St Paul, Minn) aortic valve prosthesis. The valve shows very favorable hemodynamic results even for small implants, but there are aspects that we believe should be discussed more closely.

During the study period, 15 valves had to be explanted for valve dysfunction (4 cases of nonstructural dysfunction and 11 structural valve failures), and these explantations were not further elaborated on in the text. When we implanted the Trifecta valve in our institution, we encountered several cases in which, during manipulation of the valve in the root during positioning or knotting, deformation of 1 or more struts led to distortion of the valve, which in turn led to severe central leakage. This leakage necessitated early explantation of the device, which was performed during the same session as the implantation. One might then speculate whether a smaller degree of stent distortion, not sufficient to lead to very early valve dysfunction and thus

immediate explantation, may have had an effect on a possible premature degeneration of the prosthesis. If this were the case, what would be an adequate means of detecting such distortion intraoperatively? Would it be justified to recommend explantation of these prostheses with mild central incompetence under intraoperative echocardiographic control? It would be interesting to know whether Goldman and colleagues¹ have encountered this phenomenon during their experience, particularly in patients with early degeneration.

In addition, there is another consideration in association with the implantation of these prostheses in young patients. Although this prosthesis allows a favorable hemodynamic profile even when small sizes are implanted, would it still be better to carry out an annular enlargement and use a larger implanted prosthesis to allow a subsequent valve-in-valve procedure, such as would not be possible with a size 21 or smaller valve?

Giuseppe Santarpino, MD

Steffen Pfeiffer, MD

Department of Cardiac Surgery

Paracelsus Medical University

Nuremberg, Germany

Reference

1. Goldman S, Cheung A, Bavaria JE, Petracek MR, Groh MA, Schaff HV. Midterm, multicenter clinical and hemodynamic results for the Trifecta aortic pericardial valve. *J Thorac Cardiovasc Surg.* 2017;153:561-9.e2.

<http://dx.doi.org/10.1016/j.jtcvs.2017.04.034>

BIGGER VALVE SIZE IS NOT ALWAYS BETTER

Reply to the Editor:

In their letter to the Editor, Drs Santarpino and Pfeiffer question whether early dysfunction of the Trifecta valve (St Jude Medical, St Paul, Minn) is due to the leaflets or stent and whether implantation of a small valve size should be avoided. In our study, stent deformation may have contributed to an early explant in 4 cases, whereas in 11 cases leaflet dysfunction due to calcification, fibrous thickening, or pannus formation was the primary contributing factor.¹ Although there are multiple factors that can contribute to valve dysfunction, our results demonstrate that the Trifecta valve exhibits excellent durability over 6 years of follow-up.

On the basis of our experience, the Trifecta valve is uniquely suitable for implantation in patients with a small aortic root. The best clinical outcomes occur when the valve is appropriately sized to match the aortic annulus and anatomy of the sinotubular junction. When encountering an aorta with a narrow root, attention must be given to not

The Editor welcomes submissions for possible publication in the Letters to the Editor section that consist of commentary on an article published in the Journal or other relevant issues. Authors should: • Include no more than 500 words of text, three authors, and five references. • Type with double-spacing. • See <http://jtcvs.ctsnetjournals.org/misc/fora.shtml> for detailed submission instructions. • Submit the letter electronically via jtcvs.editorialmanager.com. Letters commenting on an article published in the JTCVS will be considered if they are received within 6 weeks of the time the article was published. Authors of the article being commented on will be given an opportunity of offer a timely response (2 weeks) to the letter. Authors of letters will be notified that the letter has been received. Unpublished letters cannot be returned.