Author has nothing to disclose with regard to commercial support.

BioGlue has been extensively used in the clinical setting of type A dissection, either to promote distal false-lumen thrombosis, to secure anastomotic hemostasis, or to reinforce a dissected root. Reports of systemic emboli, pseudoaneurysm formation, and tissue necrosis linked to the possible use of BioGlue, however, have hampered its use.^{3,5,6} Despite these concerns, with appropriate use, BioGlue has been reported as a safe and useful adjunct in type A repair as suggested by Lee and Choi. Currently, there is general consensus that BioGlue should not be applied in the distal false lumen to promote thrombosis because of the risk of embolization. The use of novel hybrid covered or uncovered stent grafts is emerging as a much more effective and secure therapy to promote distal false-lumen thrombosis. The risk of BioGlue leakage through anastomosis or needle holes contraindicates its use on anastomotic surfaces. New Hemo-Seal 1:1 needles (Johnson and Johnson, New Brunswick, NJ) and still newer and safer hemostatic agents are available to enhance anastomotic hemostasis in type A dissection. Root management in type A dissection is tailored according to the extent of root destruction and includes root replacement, isolated replacement of the noncoronary sinus, or full root preservation with or without BioGlue and Teflon (The Chemours Company, Wilmington, Del) felt reinforcement. Use of BioGlue at the root level necessitates technical precautions to maximize efficacy and minimize complications. Because blood interferes with the bonding of BioGlue to the native aortic tissue, all clots within the false lumen should be removed and the aortic tissue thoroughly dried before BioGlue application. A fine layer of 1 to 2 mm of BioGlue should be spread uniformly within the false lumen. After application, the dissected membrane should be aligned and held firmly to the outer aortic wall to ensure a uniform bonding. Protection of surrounding structures with humidified sponges is important to prevent spillage of BioGlue on structures such as the aortic leaflets. Caution is advised when applying BioGlue in presence of a circular coronary ostia dissection, because BioGlue, as a result of its stiffness, has been associated with strictures when applied circumferentially in pediatric vascular structures or coronary anastomosis.

In conclusion, to use or not to use BioGlue in type A dissection? During the past decade, thoracic aortic surgery has seen the development of important adjuncts in terms of operative techniques, devices, sutures, and hemostatic agents allowing to perform extensive and safe repairs in type A dissection. Although BioGlue remains part of the “toolbox,” its use should be restricted to reinforcing the

dissected aortic root tissues. Strict adherence to the principles of safe application and knowledge of the potential limitations and complications associated with BioGlue are important to minimize the morbidity reported with its misuse.

Francois Dagenais, MD
Department of Cardiac Surgery
Institut Universitaire de Cardiologie et Pneumologie de
Québec
Québec, Québec, Canada

References

1. Yang B, Malik A, Waidley V, Kleeman KC, Wu X, Norton EL, et al. Short-term outcomes of a simple and effective approach to aortic root and arch repair in acute type A aortic dissection. *J Thorac Cardiovasc Surg*. 2018;155:1360-70.e1.
2. Witter K, Tonar Z, Matejka VM, Martinca T, Jonák M, Rokosný S, et al. Tissue reaction to three different types of tissue glues in an experimental aorta dissection model: a quantitative approach. *Histochem Cell Biol*. 2010;133:241-59.
3. Bhamidipati C, Coselli JS, LeMaire SA. BioGlue in 2011: what is its role in cardiac surgery? *J Extra Corpor Technol*. 2012;44:6-12.
4. Fürst W, Banerjee A. Release of glutaraldehyde from an albumin-glutaraldehyde tissue adhesive causes significant in vitro and in vivo toxicity. *Ann Thorac Surg*. 2005;79:1522-8; discussion 1529.
5. Kobayashi T, Kurazumi H, Sato M, Gohra H. Pseudoaneurysm rupture after acute type A dissection repair: possible reaction to BioGlue. *Interactive Cardiovasc Thorac Surg*. 2018;26:331-2.
6. Nakahira J, Ishii H, Sawai T, Minami T. Fibrin glue on an aortic cusp detected by transesophageal echocardiography after valve-sparing aortic valve replacement: a case report. *J Med Case Rep*. 2015;9:53.

<https://doi.org/10.1016/j.jtcvs.2018.10.102>

DON'T THROW OUT THE BABY WITH THE BATHWATER!

Reply to the Editor:

The letter from Lee and Choi¹ and the article by Yang and colleagues² are part of a never-ending debate about trade-offs between functionality and biocompatibility of biological glues. Lee and Choi¹ describe the use of bovine serum albumin-glutaraldehyde to repair dissected aortic sinuses, and Yang and colleagues² report the (good) results of a technique that completely avoids the use of any glue or foreign material. There is no doubt that the introduction of biological glues into clinical practice improved the early results of surgery for acute aortic dissections. But, as stated by Downing,³ “there is no free lunch at this table.”

The most powerful adhesive⁴ is cyanoacrylate, which quickly polymerizes when applied. However, the possible complications (it can cause local cell death with subsequent necrosis of the arterial wall where applied³) limited its widespread clinical application.

Another tissue adhesive is represented by a mixture of gelatin-resorcinol cross-linked with formaldehyde was

Authors have nothing to disclose with regard to commercial support.

chosen for its high bond strength even in presence of moisture. Guilmet and colleagues⁵ popularized its use in aortic dissection, even if some complications were described⁶ related to injury of the aortic tissue by formaldehyde. Lesions did not appear immediately, but in the midterm caused an elevated incidence of reoperations.^{7,8} To limit the potential toxicity of formaldehyde, bovine serum albumin-glutaraldehyde glue was introduced with the expectation of a lower toxicity. The 2 solutions mix completely within the delivery tip during application with a fixed ratio and begin to polymerize immediately, reaching full bonding strength within 2 minutes. However, the compound releases a high amount of glutaraldehyde, exposing the tissues to damages, which can include the passage of the toxic glutaraldehyde into the systemic circulation through cell-saving devices.⁹ In fact, serum albumin-glutaraldehyde, being more liquid in consistency than the gelatin-resorcinol-formaldehyde glue, has a greater chance to exude out of the false lumen.

In general, complications related to the use of biologic adhesives consist in pseudoaneurysms (similar to what can happen in patients in whom biological glues are not used) and in redissection of the aortic root that goes retrogradely toward the aortic valve,⁸ a new complication consequent to glue toxicity.

However, not all reports are negative. There is no doubt that the use of biological glues has reduced the early mortality of repair of aortic dissection, and as a consequence, the higher incidence of late complications is a low price to pay. The ratio between the compounds is fixed when serum albumin-glutaraldehyde glue is used, but can vary with gelatin-resorcinol-formaldehyde, when it is advisable to use the lower dose possible of formaldehyde (~1:10).¹⁰ Many studies do not report complications or report a complications rate similar to cases where glue was not used.^{11,12} Then, as always in surgery, we need some common sense. In conventional patients, the use of glue is not recommended except in patients in whom the fragility of tissues makes the possibility of severe bleeding likely. In aortic dissection, the use of glue can be a tool to spare the aortic root and the aortic valve, especially in low-volume centers.¹ Particular attention has to be paid to the distal anastomosis, in which the possibility of distal embolism is real.¹³⁻¹⁵ Lee and Choi,¹ even if supporting a liberal use of foreign material and biological glue to repair the dissected aortic root, do not recommend the use of glue for the repair of distal aortic dissection, and this seems a wise strategy. Biological glue is a tool to improve the results of surgery for acute aortic dissection. However, it

needs to be used carefully, taking all the precautions necessary to reduce at the minimum the amount of cross-linking agent (a toxic aldehyde) and to ensure that the area where the glue has to be applied is dry and bloodless. As always in surgery, we have only to apply the proper strategy to the patient we have on the table, using all the technical tools at our disposition without any prejudice.

Antonio Calafiore, MD^a

Antonio Totaro, MD^a

Massimiliano Foschi, MD^b

Michele Di Mauro, MD, PhD^b

^aCardiology and Cardiac Surgery

John Paul II Foundation

Campobasso, Italy

^bDepartment of Heart Disease

SS Annunziata Hospital

Chieti, Italy

References

1. Lee MK, Choi JB. Does the use of polytetra fluoroethylene felt and biological glue in the repair of acute type A aortic dissection significantly impact the surgical outcome? *J Thorac Cardiovasc Surg.* 2019;157:e118-9.
2. Yang B, Malik A, Waidley V, Kleeman KC, Wu X, Norton EL, et al. Short-term outcomes of a simple and effective approach to aortic root and arch repair in acute type A aortic dissection. *J Thorac Cardiovasc Surg.* 2018;155:1360-1367.e1.
3. Downing SW. What are the risks of using biologic glues? *Ann Thorac Surg.* 2003;75:1063. author reply 1064.
4. Carvalho MVH, Marchi E, Lourenco EA. Comparison of arterial repair through the suture, suture with fibrin or cyanoacrylate adhesive in ex-vivo porcine aortic segment. *Braz J Cardiovasc Surg.* 2017;32:487-91.
5. Guilmet D, Bachet J, Goudot B, Laurian C, Gigou F, Bical O, et al. Use of biological glue in acute aortic dissection. Preliminary clinical results with a new surgical technique. *J Thorac Cardiovasc Surg.* 1979;77:516-21.
6. Hata H, Takano H, Matsumiya G, Fukushima N, Kawaguchi N, Sawa Y. Late complications of gelatin-resorcin-formalin glue in the repair of acute type A aortic dissection. *Ann Thorac Surg.* 2007;83:1621-6.
7. Suzuki S, Imoto K, Uchida K, Takanashi Y. Aortic root necrosis after surgical treatment using gelatin-resorcinol-formaldehyde (GRF) glue in patients with acute type A aortic dissection. *Ann Thorac Cardiovasc Surg.* 2006;12:333-40.
8. Nakajima T, Kawazoe K, Izumoto H, Kamada T, Kataoka T, Yoshioka K, et al. Influence of gelatin-resorcin-formalin glue on mid-term redissection after aortic repair. *Surg Today.* 2005;35:112-6.
9. Furst W, Banerjee A. Release of glutaraldehyde from an albumin-glutaraldehyde tissue adhesive causes significant in vitro and in vivo toxicity. *Ann Thorac Surg.* 2005;79:1522-9.
10. Kuniyara T, Shiya N, Matsuzaki K, Murashita T, Matsui Y. Recommendation for appropriate use of GRF glue in the operation for acute aortic dissection. *Ann Thorac Cardiovasc Surg.* 2008;14:88-95.
11. Yamanaka K, Hori Y, Ikarashi J, Kusuhara T, Nakatsuka D, Hirose K, et al. Durability of aortic valve preservation with root reconstruction for acute type A aortic dissection. *Eur J Cardiothorac Surg.* 2012;41:e32-6.
12. Ma WG, Ziganshin BA, Guo CF, Zafar MA, Sieller RS, Tranquilli M, et al. Does BioGlue contribute to anastomotic pseudoaneurysm after thoracic aortic surgery? *J Thorac Dis.* 2017;9:2491-7.
13. Bernabeu E, Castella M, Barriuso C, Mulet J. Acute limb ischemia due to embolization of biological glue after repair of type A aortic dissection. *Interact Cardiovasc Thorac Surg.* 2005;4:329-31.
14. Carrel T, Maurer M, Tkebuchava T, Niederhauser U, Schneider J, Turina MI. Embolization of biologic glue during repair of aortic dissection. *Ann Thorac Surg.* 1995;60:1118-20.
15. Furukawa H, Masaki H, Tanemoto K. Late limb embolization of biological glue after repair of aortic dissection. *Eur J Cardiothorac Surg.* 2015;48:633.

<https://doi.org/10.1016/j.jtcvs.2018.11.008>