particularly women. Other studies, particularly with transcatheter aortic valve replacement, have demonstrated why the investigation of sex-based differences in procedural outcomes is critical.^{3,4} Although the first publication of results with the Relay device featured gender parity with 52.6% being men, of the 26 patients who required iliac conduits in that study, 21 (80.8%) were women, with 1 woman ultimately being deemed anatomically unsuitable.² Conduits were only required in 2 patients (2.9%) with the RelayPro and technical success was achieved in 100% of patients.

This all goes to show that the constant evolution of medical technologies is not only important for the goal of continuous improvement in clinical outcomes in general, but also for the principle of providing high-quality care to

the wide variety of patients that we treat, regardless of sex, age, or race (see [Figure 1](#)).

References

1. Szeto WY, Vallabhajosyula P, Matsuda H, Moainie SL, Sharafuddin MJ, Corvera J, et al. One-year results with a low-profile endograft in subjects with thoracic aortic aneurysm and ulcer pathologies. *J Thorac Cardiovasc Surg*. 2022;163:1739-50.e4.
2. Farber MA, Lee WA, Szeto WY, Panneton JM, Kwolek CJ. Initial and midterm results of the Bolton Relay Thoracic Aortic Endovascular Pivotal Trial. *J Vasc Surg*. 2017;65:1556-66.e1. <https://doi.org/10.1016/j.jvs.2016.11.061>
3. Chandrasekhar J, Dangas G, Yu J, Vemulapalli S, Suchindran S, Bora AN, et al. Sex-based differences in outcomes with transcatheter aortic valve therapy: TVT Registry from 2011 to 2014. *J Am Coll Cardiol*. 2016;68:2733-44. <https://doi.org/10.1016/j.jacc.2016.10.041>
4. Williams M, Kodali SK, Hahn RT, Humphries KH, Nkomo VT, Cohen DJ, et al. Sex-related differences in outcomes after transcatheter or surgical aortic valve replacement in patients with severe aortic stenosis: insights from the partner trial (placement of aortic transcatheter valve). *J Am Coll Cardiol*. 2014;63:1522-8. <https://doi.org/10.1016/j.jacc.2014.01.036>

See Article page 1739.

Commentary: Profile matters: 1-year results of low profile endograft with thoracic aortic aneurysm and ulcer pathologies

Sashi K. Inkollu, MD,^a and
Carlos F. Bechara, MD, DFSVS^b

Sashi K. Inkollu, Vascular Surgeon, MCHS, and Carlos F. Bechara, Professor, LUMC

Some of the most common complications encountered in treating thoracic aortic pathologies are related to vascular access such as bleeding from failure of percutaneous access closure devices and iliac rupture as

CENTRAL MESSAGE

The newer-generation RelayPro thoracic device (Terumo Corp) offers a lower profile and better conformability but will its long-term results hold good for device integrity?

From the ^aDepartment of Vascular Surgery, Marshfield Clinic Health Systems, Marshfield, Wis; and ^bDepartment of Vascular Surgery, Loyola University Medical Center, Maywood, Ill.

Disclosures: Dr Bechara is a consultant and speaker for Gore Medical, Cook Medical, Terarecon, and Penumbra, owns stock in Mokita Medical, and is a consultant (CEC) for Endospa. Dr Inkollu reported no conflicts of interest.

The *Journal* policy requires editors and reviewers to disclose conflicts of interest and to decline handling or reviewing manuscripts for which they may have a conflict of interest. The editors and reviewers of this article have no conflicts of interest.

Received for publication Nov 16, 2021; revisions received Nov 16, 2021; accepted for publication Nov 17, 2021; available ahead of print Nov 20, 2021.

Address for reprints: Sashi K. Inkollu, MD, Department of Vascular Surgery, Marshfield Clinic Health Systems, 1000 N Oak Ave, Suite 3C, Marshfield, WI 54449 (E-mail: inkollu.sashi@marshfieldclinic.org).

J Thorac Cardiovasc Surg 2022;163:1752-3

0022-5223/\$36.00

Copyright © 2021 by The American Association for Thoracic Surgery

<https://doi.org/10.1016/j.jtcvs.2021.11.030>

well as access vessel occlusion from dissection and thrombosis. These complications are more common in the presence of smaller-access vessels relative to the device profile such as in women or in the presence of occlusive disease. Szeto and colleagues¹ have presented the 1-year results of the new-generation low profile RelayPro thoracic aortic stent (Terumo Corp) graft in the treatment of aneurysm and ulcer pathologies. This was a prospective, multicenter and international study of

110 patients with 39% Asian and 37% female participants, who typically tend to have smaller access vessels. The second-generation device is 3-4 French smaller compared with the previous generation with a technical advantage in navigating smaller access vessels apart from a shorter longitudinal curved nitinol wire that might have a better conformation to arch angulation and aortic curves. The newer-generation device compares well with the previous generation with significant improvement in major adverse events (6.4% vs 20%) and 100% technical success. Apart from smoother delivery of the device, lower profile and conformability can also reduce arch manipulation and the risk of subsequent stroke. The second-generation RelayPro device has a non-bare stent version that might have a further advantage in reducing the stroke risk by eliminating the bare stent across the arch vessels.

Reduction in the profile often raises concerns for device integrity. Although the stent diameter and design was not modified in the new generation, tightly woven polyester fabric was used in reducing the profile.

Although we all welcome lower profile devices to avoid access-related complications, we hope it will not be at the expense of device integrity as learned from other devices mentioned in the report by the authors. Hence, we urge the authors to report their long-term data at 3, 5, and 10 or more years by creating a registry. Also, 2 important aspects of this 1-year study outcomes are worth highlighting. Despite the lower profile device the cohort still had a 5.5% vascular access complications rate despite being performed by experienced surgeons. Also the primary effectiveness at 12 months was 89.2%. Despite meeting study design parameters, it highlights the importance of patient selection because these complications will stay with us despite advances of stent graft design and lower profile.

Reference

1. Szeto WY, Vallabhajosyula P, Matsuda H, Moainie SL, Sharafuddin MJ, Corvera J, et al. One-year results with a low-profile endograft in subjects with thoracic aortic aneurysm and ulcer pathologies. *J Thorac Cardiovasc Surg.* 2022;163:1739-50.e4.