LETTERS TO THE EDITOR

Perioperative use of carbon dioxide production in cardiac surgery

To the Editor:

In the November 1994 issue of the JOURNAL, Feng and Singh¹ reported the use of end-tidal carbon dioxide tension as an index of cardiac output. They monitored end-tidal carbon dioxide to facilitate weaning from cardiopulmonary bypass in adults having cardiac operations.

This approach seems to be quite reductive, because a more extensive evaluation can be obtained by adding the following (generally available) metabolic and ventilatory parameters: arterial oxygen tension, arterial carbon dioxide tension, mixed expiratory partial pressure of carbon dioxide, arterial-alveolar carbon dioxide gradient (obtained from arterial carbon dioxide tension and end-tidal carbon dioxide), alveolar-arterial oxygen gradient (evaluated with an inspired oxygen tension of 0.5), oxygen consumption, carbon dioxide production, compliance per kilogram of body weight (from the ventilator measurements tidal volume/plateau pressure in milliliters per centimeter of water per kilogram of body weight), and rest energy expenditure ($REE = \text{carbon dioxide production} \times 5.52$). All these parameters can be monitored both during the operation and during the postoperative period, to obtain a more comprehensive evaluation of the metabolic, ventilatory, and circulatory situations.

Literature reports exist on this type of monitoring, particularly with regard to postoperative evaluation in adult cardiac surgery.^{2,3} In pediatric cardiac surgery, the aforementioned methods of monitoring have been used to differentiate among the perioperative variations after correction of congenital heart defects with preoperative left-to-right shunt, with right-to-left shunt, or without intracardiac shunt. In fact, such methods have been reported in this JOURNAL.⁴

Therefore, monitoring of end-tidal carbon dioxide tension not only should be used as a routine evaluation in adult and pediatric cardiac surgery, but also should be applied to more complex situations (e.g., assisted mechanical circulation, unusual palliative procedures) to expand the horizon of its potential use.

Antonio Corno, MD
Institut Cardiovasculaire Paris Sud
66 rue du Colonel Fabien
92160 Antony, France

REFERENCES

1. Feng WC, Singh AK. Intraoperative use of end-tidal carbon dioxide tension to assess cardiac output. J THORAC CARDIOVASC SURG 1994;108:991-2.

2. Chiara O, Giomarelli P, Biagioli B, Rosi R, Gattinoni L. Hypermetabolic response after hypothermic cardiopulmonary bypass. Crit Care Med 1987;15:1071-5.
3. Corno A, Giomarelli P, Biagioli B, Barbini P, Grossi A. Postoperative cardiopulmonary monitoring with a computer-based system. J Am Coll Cardiol 1990;15:260A.
4. Giomarelli P, Biagioli B, Lisi G, Santoro F, Corno A. On-line metabolic and ventilatory monitoring in pediatric cardiac operations. J THORAC CARDIOVASC SURG 1989;97:939-40.

12/8/65889

Reply to the Editor:

We concur with Dr. Corno's opinion that a more inductive use of end-tidal carbon dioxide tension is both practical and valuable. As stated in our letter, end-tidal carbon dioxide tension is governed by metabolism, ventilation, and circulation. Therefore, it could be used to evaluate the status of any of these three values. However, we would like to reiterate the necessary reductive nature of our use of end-tidal carbon dioxide tension to estimate pulmonary flow. It cannot be over-emphasized that such an assessment will be valuable only when there is no significant deviation from the usual state of carbon dioxide production, of ventilation, and of pulmonary dead space. Furthermore, its value is time-limited, inasmuch as a low flow state will eventually raise arterial carbon dioxide levels and thus end-tidal carbon dioxide tension. As a result, we propose that this method be used only during the weaning phase of cardiopulmonary bypass as an adjunct to assess flow status, because the aforementioned variables are most likely to be stable. At the same time, we cannot fail to appreciate the simplicity and economy of such a monitoring system.

William C. Feng, MD, DSc
Arun K. Singh, MD, FRCS(C)
Division of Cardio-Thoracic Surgery
Rhode Island Hospital
Brown University School of Medicine
Providence, RI 02903
12/8/65890