- tion of the hemodynamic performance of small CarboMedics aortic prostheses using dobutamine-stress Doppler echocardiography. *Ann Thorac Surg* 1995;60:1048-52.
5. Izzat MB, Birdi I, Wilde P, Bryan AJ, Angelini GD. Comparison of the hemodynamic performance of St. Jude Medical and CarboMedics 21 mm prostheses using dobutamine stress echocardiography. *J Thorac Cardiovasc Surg* 1996;111:408-15.
 6. Kadir I, Izzat MB, Wilde P, Reeves B, Bryan AJ, Angelini GD. Dynamic evaluation of the 21 mm Medtronic Intact aortic bioprosthesis by dobutamine echocardiography. *Ann Thorac Surg*. In press.
 7. Teoh KH, Fulop JC, Weisel RD, et al. Aortic valve replacement with a small prosthesis. *Circulation* 1987;76(Suppl): III123-31.

12/8/80409

Reply to the Editor:

In their letter, Izzat and Yim essentially voice four criticisms of our results: first, that the transvalvular pressure drops that we report for mechanical prostheses are too high; second, that we should have taken measurements during exercise or after dobutamine infusion; third, that we should have investigated the dependence of patient-prosthesis mismatch on cardiac output rather than body surface area; and fourth (and presumably as a corollary of the first and third criticisms), that our recommendation that 19 mm valves not be implanted in patients with body surface areas greater than 1.7 m² is irrelevant.

Their first criticism is based on three points: a theoretical argument relating to the high-velocity central jet of bileaflet valves; the fact that we find no significant difference between the transvalvular pressure drops of bileaflet valves and bioprostheses (which they believe to have higher transvalvular pressure drops); and comparison with the transvalvular pressure drops reported in three papers authored by, among others, Izzat.

With regard to the first argument, we point out that, regardless of what the velocities of the lateral jets may be, that of the central jet is the result of a pressure that is a valid measure of ventricular effort and can be accurately calculated from the measured velocities (the value so calculated essentially does coincide with catheter measurements¹). With regard to the second point, we note that bioprostheses do not necessarily have similar hemodynamics^{2,3}; in all the small bioprostheses in our study the flaps were mounted outside the stent, and their hemodynamic performance is better than that of tilting disc valves and at least as good as that of bileaflet valves.¹⁻⁶ With regard to the third point, we frankly do not understand the transvalvular pressure drops reported in the papers they cite^{7,8} (e.g., mean transvalvular pressure drops of 3.12 ± 3.6 mm Hg for nine 21 mm St. Jude Medical valves [St. Jude Medical, Inc., St. Paul, Minn.], 4.87 ± 3.8 mm Hg for ten 21 mm CarboMedics valves [CarboMedics, Inc., Austin, Tex.], and 8.1 ± 8.4 mm Hg for eight 19 mm CarboMedics valves), which are between one half and one sixth of the values habitually found, not only by us, but by numerous other groups. For example, De Paulis and associates⁴ reported values of 20.1 ± 7.1 and 12.3 ± 3.4 mm Hg, respectively, for 19 and 21 mm CarboMedics valves; Franzen and colleagues,⁵ 17.0 ± 5.6 and 18.0 ± 7.7 mm Hg, respectively, for 21 mm CarboMedics and St.

Jude Medical valves; and Ihlen,¹ Chambers,⁶ and their coworkers, 17.1 ± 5.6 and 19.4 ± 5.6 mm Hg, respectively, for 19 mm CarboMedics valves.

In response to their second criticism, we have recently submitted for publication the exercise results obtained in the study that was mentioned in our article as still in progress. We do not believe that publication of the "disturbing" results obtained at rest with 19 mm valves should necessarily have awaited completion of the fuller study.

In response to their third criticism, transvalvular pressure drops are influenced by many interrelated factors, including valve area, systolic volume (and hence cardiac output), outflow tract anatomy, ventricular function, and body surface area. No correlation between transvalvular pressure drop and any single factor appears to be stable under variation of other factors. For example, in one of the papers by Izzat and others,⁸ there was a correlation between transvalvular pressure drop and cardiac output and transvalvular pressure drop and body surface area for one prosthesis type but not for another; unpublished results of our own, for 19 mm valves but not for larger sizes, exhibit correlations of transvalvular pressure drop with body surface area both at rest and after exercise and correlation with cardiac output only after exercise. On the other hand, quite an uncontroversial correlation exists between body surface area and aortic root size, which is what determines the size of prosthesis that can be implanted without a root enlargement procedure.

In view of the foregoing, we believe it is unnecessary to reply explicitly to their fourth criticism.

We incidentally point out what must surely be an error in two of Izzat's papers,^{7,8} both of which report a cardiac output of only 2.8 L/min in their group of 10 patients with 21 mm CarboMedics valves (of the value of 4.1 L/min they report for their 19 mm group). We also take this opportunity to point out another unconvincing feature of the data reported in one of the papers by Izzat and others⁷: the finding that whereas the effective area of 19 mm CarboMedics valves was 1.37 ± 0.53 cm², that of 21 mm valves of the same type was only 1.20 ± 0.62 cm². The problem there is not so much that the order of the mean values is the reverse of what it should logically be, but that this inversion is presumably the result of standard deviations that are much larger than those usually reported (0.1 to 0.2 cm² for 19 mm valves¹⁻⁶). One wonders whether these large standard deviations really reflect mainly inter-patient variation, especially inasmuch as no data on evaluator variability are given.

To close, we reiterate our reservations as to the routine use of 19 mm prostheses and bioprostheses. The exercise results to which Izzat and Yim look forward, and whose recent submission for publication was mentioned earlier, show that the increase in transvalvular pressure drops across small valves during effort is, to say the least, alarming.

Josè Ramòn González-Juanatey, MD, PhD
Department of Cardiology
Galician General Hospital
University of Santiago
Faculty of Medicine
Santiago de Compostela, Spain

REFERENCES

1. Ihlen H, Molstad P, Simonseu S, et al. Hemodynamic evaluation of the CarboMedics prosthetic heart valve in the aortic position: comparison of noninvasive and invasive techniques. *Am Heart J* 1992;123:151-9.
2. González-Juanatey JR, García-Acuña JM, Vega M, et al. Hemodynamics of various designs of 19 mm pericardial aortic valve bioprosthesis. *Eur J Cardiothorac Surg* 1996;10:201-6.
3. González-Juanatey JR, García-Bengoechea JB, García-Acuña JM, et al. The influence of the design on the medium to long term hemodynamic behavior of 19 mm pericardial aortic valve prostheses. *J Heart Valve Dis* 1996;5(Suppl III):S317-S23.
4. De Paulis R, Sommariva L, Russo F, et al. Doppler echocardiographic evaluation of the CarboMedics valve in patients with small aortic anulus and valve prosthesis-body surface area mismatch. *J Thorac Cardiovasc Surg* 1994;108:57-62.
5. Franzen SF, Hultjebrant IE, Konstantinov IE, Nylander E, Olin CL. Aortic valve replacement for aortic stenosis in patients with small aortic root. *J Heart Valve Dis* 1996;5(Suppl III):S284-S8.
6. Chambers J, Cross J, Deverall P, Sowton E. Echocardiographic description of the CarboMedics bileaflet prosthetic heart valve. *J Am Coll Cardiol* 1993;21:398-405.
7. Izzat MD, Birdi Y, Wilde P, Bryan AJ, Angelini GD. Evaluation of the hemodynamic performance of small CarboMedics aortic prostheses using dobutamine-stress Doppler echocardiography. *Ann Thorac Surg* 1995;60:1048-52.
8. Izzat MD, Birdi Y, Wilde P, Bryan AJ, Angelini GD. Comparison of the hemodynamic performance of the St. Jude Medical and CarboMedics 21 mm prostheses using dobutamine stress echocardiography. *J Thorac Cardiovasc Surg* 1996;111:408-15.

12/8/80407

Myocardial protection of the hypoxic heart

To the Editor:

The study of Bolling and associates¹ on myocardial protection in normal and hypoxically stressed hearts in this Journal concluded that good myocardial protection is independent of cardioplegic calcium concentration in nonhypoxic hearts and that hypoxic hearts are extremely sensitive to the cardioplegic calcium concentration. The authors lamented the absence of a clinically relevant hypoxic model. This is an important consideration, because many children who undergo cardiac operations in the first year of life have cyanotic heart disease, and the myocardium may be chronically perfused with hypoxic blood. Clearly it would be desirable to alter current techniques of myocardial management during surgery to provide better myocardial protection for children with cyanotic congenital heart disease and to further determine whether such techniques might also be applied to adults with acyanotic acquired heart disease. If the impact of prolonged periods of hypoxia on tolerance to subsequent ischemia could be elucidated, cardioprotection in these patients could be better understood and improved. The model of acute hypoxia described by Bolling and colleagues,¹ which was originally described by Buckberg,² Ihnken,³ and their associates, does not allow for the development of adaptive mechanisms in response to

chronic hypoxia that occur in children with cyanotic heart disease. In addition, it is unknown whether reoxygenation is a real source of cardiac dysfunction in the chronically hypoxic immature heart. Preexisting injury caused by acute hypoxia may be a prerequisite for reoxygenation injury in the studies described by Bolling's group.¹

Our laboratory has developed a rabbit model of hypoxia from birth that simulates the essential characteristics of cyanotic heart disease: decreased arterial oxygen levels, polycythemia, increased right ventricular mass, decreased weight gain, and overall failure to thrive.⁴ We have demonstrated that hypoxia from birth increases the tolerance of the heart to ischemia compared with that of age-matched normoxic control subjects.⁴ The mechanism for adaptation to prolonged hypoxia that confers increased tolerance to subsequent myocardial ischemia is currently unknown.

We have also examined the calcium content of St. Thomas' Hospital II cardioplegic solution and found in our model that the existing calcium content of 1.2 mmol/L is suboptimal to protect the ischemic immature myocardium. Optimal myocardial protection occurred with a calcium concentration of 0.4 mmol/L in hearts chronically hypoxic from birth⁵ and 0.3 mmol/L in hearts normoxic from birth.⁶ These previously published data may place in perspective the study by Bolling and coworkers, confirming their hypothesis of hypocalcemic cardioplegia being superior to normocalcemic cardioplegia in protection of immature myocardium. The response, however, of immature myocardium to acute versus chronic hypoxia before reoxygenation, with and without surgical ischemia, is speculative because neither we nor Bolling and his associates addressed this particular issue.

John E. Baker, PhD

Gordon N. Olinger, MD

Department of Cardiothoracic Surgery

Medical College of Wisconsin

Milwaukee, WI 53226

REFERENCES

1. Bolling K, Kronon M, Allen BS, Ramon S, Wang T, Hartz RS, et al. Myocardial protection in normal and hypoxically stressed neonatal hearts: the superiority of hypocalcemic versus normocalcemic blood cardioplegia. *J Thorac Cardiovasc Surg* 1996;112:1193-201.
2. Buckberg GD. Studies of hypoxemic/reoxygenation injury. I. Linkage between cardiac function and oxidant damage. *J Thorac Cardiovasc Surg* 1995;110:1164-70.
3. Ihnken K, Morita K, Buckberg GD, Matheis G, Sherman MP, Allen BS, et al. Studies of hypoxemic/reoxygenation injury. II. Evidence for reoxygenation damage. *J Thorac Cardiovasc Surg* 1995;110:1171-81.
4. Baker EJ, Boerboom LE, Olinger GN, Baker JE. Tolerance of the developing heart to ischemia: impact of hypoxemia from birth. *Am J Physiol* 1995;268:H1165-73.
5. Baker EJ, Baker JE. Calcium and cardioplegic protection of the ischemic immature heart: impact of hypoxemia from birth. *Ann Thorac Surg* 1994;58:1123-30.
6. Baker EJ, Olinger GN, Baker JE. Calcium content of St. Thomas' II cardioplegic solution damages ischemic immature myocardium. *Ann Thorac Surg* 1991;52:993-9.

12/8/80740