Fig 1. The BioCor stentless mitral valve. It is apparent how short the chordae are. Since the length of these chordae is fixed and limited, mismatch can occur leading to chordal rupture or stiffening. The relatively large amount of pericardial tissue used to strengthen the chordae causes further limitation in the movement of the cusps.

more, if this mitral stentless valve is implanted in a child and the heart grows, the chordae are pulled down toward the ventricle, preventing the normal movement of the cusps. In both cases, the mitral chordae ruptured because of undue tension. Nonmoving cusps will eventually calcify as well. Interestingly, in 1 child the aortic valve had a minimal amount of calcification, and the second child had a normal valve 3 years after the operation. The valve was not explanted and is still functioning well years after the operation.

I differ with Milo and associates about the causes of calcification of the BioCor mitral valve treated with the No-React formula. The BioCor stentless mitral valve is an unproved design. Its implantation and proper function depend on many factors, which can cause the valve to stop moving or even tear without any degeneration. The fact that these patients had severe endocarditis could also be a factor that caused inflammation or subclinical endocarditis followed obviously by calcification. Clinical experience with BioCor stentless valves treated with the No-React formula so far shows that these valves resist calcification. Von Oppell and coworkers¹ recently presented an experience with the BioCor No-React stentless aortic valve in 52 consecutive patients. The mean age was 44 ± 17 years, and the youngest patient was 16 years at the time of implantation. The follow-up period is already 5 years (first implant in 1994). Although this population is considered to be very young for a bioprosthesis, no calcification has been seen yet.

Clinical experience with the pulmonic porcine valve conduit treated with the No-React formula has been reported in infants younger than 1 year, with a follow-up period of 18 months. No calcification has been seen so far.² In these young children calcification of homografts can be seen as early as 4 months after implantation.³ The fact that the aortic stentless valve treated with the No-React formula resisted calcification whereas the BioCor mitral stentless valve did not, in the same patients, gives further credence to the claim that anticalcification treatment cannot compensate for nonfunctioning or

infected valves. I believe that when there is a need to implant a bioprosthesis in the mitral position in young patients, it should be a proven design.

The stentless mitral xenograft is a complex valve that needs more evaluation. It should not be recommended for use in children. The conclusions of Milo and associates I believe should concentrate on the difference between the aortic and mitral stentless valves in the presence of endocarditis. After all, if the native valve calcifies after an infection, it will be presumptuous to expect any anticalcification treatment to be superior to the patient's own native tissue.

Shlomo Gabbay, MD
Department of Cardiothoracic Surgery
University of Medicine and Dentistry of New Jersey
New Jersey Medical School
185 South Orange Ave
Newark, NJ 07103-2757

REFERENCES

1. Von Oppell U, Stemmet F, Brink J, LeVetan B, Heijke S. BioCor No-React stentless aortic valve: short term results. Presented at the Stentless Bioprostheses Third International Symposium. 1999; May 12-16 (abstract).
2. Massetti P, Ussia GP, Gazzolo D, Marianeschi MS, Abella FR, Cipriani A, et al. Aortic pulmonary autograft implant: medium term follow-up with a note on a new right ventricular pulmonary artery conduit. *J Card Surg* 1998;13:173-6.
3. Clarke RD, Bishop AD. Ten year experience with pulmonary allografts in children. *J Heart Valve Dis* 1995;4:381-4.

12/8/100891

Reply to the Editor:

Dr Gabbay was deeply involved in choosing the type of stentless valves used in these 2 children and in observing them after the operation. We informed him of the results of all echocardiographic studies done on the children. Early post-operative studies in the 2 children showed that both leaflets of the stentless mitral valves had a full spectrum of movement and that both coaptation and closure were good. In fact, an inexperienced observer could barely recognize an implanted prosthesis in the mitral position. Furthermore, one of the authors (Dr Bar-El) was invited to present these promising early results at the 1996 meeting of The Society of Thoracic Surgeons at Gabbay's presentation on the No-React detoxification process (Shelhigh Inc, Millburn, NJ).¹ One month later, as the children's hearts began to recover and return to near normal size, we noted a progressive prolapse of both leaflets. As the prolapse steadily increased, we began to consider replacing the valves, not because of the length of the chordae tendineae, but because of the prolapse itself. However, calcification, not the prolapse, determined the time of surgical intervention. Our follow-up echocardiographic data do not support Gabbay's contention that torn chordae resulted from undue tension. We believe that the early biodegradation of the valve resulted in calcification, which led to rupture of the chordae.

We do believe that a stentless mitral prosthesis that incorporates the entire subvalvular apparatus may cause problems

Fig 1. Cusps of explanted BioCor stentless aortic valve treated with the No-React formula, which was implanted 5 years ago. Note disintegration of 2 of the cusps.

when implanted in hearts that may eventually change size. Our results with the No-React-treated BioCor stented and stentless valves (BioCor, Belohorizonte, Brazil) in adults correlate with those of Von Oppell and associates² but are not different from results obtained with valves in which the glutaraldehyde was not rinsed out. Massetti and associates³ reported that No-React pulmonic porcine valve conduits they implanted in children were free from calcification at 18 months, and we are hopeful that the valves will remain so for several years. In our study, the valves were free from calcification at 18 months too. However, on June 11, 1999, we had to explant the last stentless BioCor No-React-treated aortic valve on an emergency basis because of rupture of 2 of its cusps (Fig 1). The left and right coronary cusps had entirely disintegrated at the free edges, but we saw no evidence of calcification. The condition of the valve caused acute and severe aortic incompetence. We may learn about the need to detoxify the tissues and the potential effect on its mechanical properties from this last patient.

Simcha Milo, MD
Department of Cardiac Surgery
Rambam Medical Center
Haifa 35254, Israel

REFERENCES

1. Abolhoda A, Yu S, Oyarzun JR, et al. No-React detoxification process: a superior anticalcification method for bioprostheses. *Ann Thorac Surg* 1996;62:1724-30.
2. Von Oppell U, Stemmet F, Brink J, LeVetan B, Heijke S. BioCor

No-React stentless aortic valve: short term results. Presented at the Stentless Bioprostheses Third International Symposium. 1999; May 12-16 (abstract).

3. Massetti P, Ussia GP, Gazzolo D, Marianeschi MS, Abella FR, Cipriani A, et al. Aortic pulmonary autograft implant: medium term follow-up with a note on a new right ventricular pulmonary artery conduit. *J Card Surg* 1998;13:173-6.

12/8/100892

Percutaneous treatment of acute aortic dissection

To the Editor:

Acute aortic dissection is frequently a surgical emergency. The recent publications of nonsurgical treatment with catheter-delivered endovascular stent-grafts by the Stanford group^{1,2} and the German-Italian group³ cast a new light on the approach to this life-threatening disease. These studies represent a milestone in the treatment of aortic dissection.

Cardiovascular catheterization and angiography are no longer for diagnosis only. They have blossomed into therapeutic applications and become an integral part of therapeutic interventions in both cardiac and vascular diseases. The Stanford experiences also represent a multi-disciplinary effort involving both cardiovascular surgeons and interventional radiologists. Stent-graft placement requires state-of-the-art image technology, as well as precise manipulation of the catheter and positioning of the stent.⁴ The results are very promising and the future looks bright. We await more patients to receive such innovative percutaneous treatment and their long-term results.

Tsung O. Cheng, MD
Professor of Medicine
The George Washington University
Washington, DC 20037

REFERENCES

1. Slonim SM, Miller DC, Mitchell RS, Semba CP, Razavi MK, Dake MD. Percutaneous balloon fenestration and stenting for life-threatening ischemic complications in patients with acute aortic dissection. *J Thorac Cardiovasc Surg* 1999;117:1118-27.
2. Dake MD, Kato N, Mitchell RS, Semba CP, Razavi MK, Shimono T, et al. Endovascular stent-graft placement for the treatment of acute aortic dissection. *N Engl J Med* 1999;340:1546-52.
3. Nienaber CA, Fattori R, Lund G, Dieckmann C, Wolf W, von Kodolitsch Y, et al. Nonsurgical reconstruction of thoracic aortic dissection by stent-graft placement. *N Engl J Med* 1999;340:1539-45.
4. Vlahakes GJ. Catheter-based treatment of aortic dissection (editorial). *N Engl J Med* 1999;340:1585-6.

12/8/101203