

The diameters of the pig tracheas in this study (range, 11-14 mm) are larger than human neonatal and infant airways that may require intervention before age 1 year (range, 5-6 mm). Further dynamic longitudinal studies in smaller animals may be needed to determine whether the stent accomplishes its intended goal of preventing pediatric airway collapse.

These preclinical results show promise for a new type of stent with successful initial performance in an animal airway model. The potential influence of this new device is excellent, and it could fill a critical vacancy where no good current options exist. This represents another example of important surgical innovation from an

exemplary team, and further clinical progress is eagerly anticipated.

References

1. Kamran A, Jennings RW. Tracheomalacia and tracheobronchomalacia in pediatrics: an overview of evaluation, medical management, and surgical treatment. *Front Pediatr.* 2019;7:512.
2. Valerie EP, Durrant AC, Forte V, Wales P, Chait P, Kim PC. A decade of using intraluminal tracheal/bronchial stents in the management of tracheomalacia and/or bronchomalacia: is it better than aortopexy? *J Pediatr Surg.* 2005;40:904-7.
3. Mondal A, Ha J, Jo VY, Wu F, Kaza AK, Dupont PE. Preclinical evaluation of a pediatric airway stent for tracheobronchomalacia. *J Thorac Cardiovasc Surg.* 2021;161:e51-60.


See Article page e51.



Commentary: Toward a more ideal pediatric airway stent for tracheobronchomalacia

Roosevelt Bryant III, MD

The first clinical deployment of a pediatric airway stent was in 1988 and was reported by Loeff and colleagues.² Since that time, an array of airway stents have developed to treat complex airway disease in children. Pediatric-specific applications of airway stents include use after tracheal reconstruction for congenital tracheal stenosis³ and for tracheobronchomalacia⁴ not responsive to medical therapy. However, the ideal pediatric airway stent has yet to be developed. The ideal airway stent for pediatric patients should be easy to place, should support the airway without the development of significant complications, and should be easy to remove to allow maximal growth of the airway.³



CasMin Twine helical stent.¹

CENTRAL MESSAGE

The helical Niti-S airway stent shows promise as a more ideal prosthesis for the management of tracheobronchomalacia.

From the Division of Congenital Heart Surgery, The Heart Center, Phoenix Children’s Hospital, Phoenix, Ariz.

Disclosures: The author reported no conflicts of interest.

The *Journal* policy requires editors and reviewers to disclose conflicts of interest and to decline handling or reviewing manuscripts for which they may have a conflict of interest. The editors and reviewers of this article have no conflicts of interest.

Received for publication April 3, 2020; revisions received April 3, 2020; accepted for publication April 3, 2020; available ahead of print April 18, 2020.

Address for reprints: Roosevelt Bryant III, MD, Heart Transplantation and Mechanical Circulatory Support, Phoenix Children’s Hospital, 1919 E Thomas Rd, Phoenix, AZ 85016 (E-mail: rbryant1@phoenixchildrens.com).

J Thorac Cardiovasc Surg 2021;161:e62-3

0022-5223/\$36.00

Copyright © 2020 by The American Association for Thoracic Surgery

<https://doi.org/10.1016/j.jtcvs.2020.04.035>

Potential stent-related complications include migration, granulation tissue formation, mucus formation, and infection,⁵ particularly when granulation tissue develops. Despite the wide array (metallic,⁶ silicon,⁷ bioabsorbable⁸) of pediatric stents available, none of them is ideal. The known complications associated with these devices have led to a stent-related mortality rate as high as 12.9%.⁷ Furthermore, the radial force used to keep certain kinds of stents in place has been shown to damage the microcirculation and serves as the nidus for mucosal injury and subsequent granulation tissue formation.^{9,10}

In this issue of the *Journal*, Mondal and colleagues¹¹ present the results of a promising animal study looking at a novel airway stent, the helical Niti-S stent (Taewong Medical, Seoul, Korea), for the treatment of tracheobronchomalacia in children.¹¹ Its helical design is similar to the CasMin-Twine helical stent¹ used to treat canine tracheal collapse. The 2 types of helical stents share certain advantages. These include ease of deployment, limited mucus production, and focal mucosal changes limited to the areas of stent contact. The helical design provides radial support with substantially less mucosal contact and at a lower pressure. The results of the study by Mondal colleagues provide further evidence that the stent design provides less of a nidus for granulation tissue formation. This was also shown with the CasMin-Twine helical stent animal study. Where the 2 stents diverge is the novel removal tool associated with the helical Niti-S stent. Removal of the stent should allow optimal growth of the airway. A primary limitation of the study by Mondal colleagues is the short duration of therapy. After tracheal reconstruction for example, it can take 2 to 6 months for the airway to endothelialize.³ It remains to be seen whether the helical Niti-S stent will be as amenable to removal with longer durations of therapy. However, the early advantages of the helical Niti-S certainly seem to be taking us toward a more ideal pediatric airway stent.

References

1. Lopez-Minguez S, Serrano-Casorran C, Guirola JA, Rodriguez-Zapater S, Bonastre C, DeGregorio MA. New tracheal stainless steel pilot study: twelve month follow-up in a rabbit model. *PeerJ*. 2019;7:e7797.
2. Loeff DS, Filler RM, Gorenstein A, Ein S, Philippart A, Bahoric A, et al. A new intratracheal stent for tracheobronchial reconstruction: experimental and clinical studies. *J Pediatr Surg*. 1988;23:1173-7.
3. Jacobs JP, Quintessenza JA, Botero LM, van Gelder HM, Giroud JM, Elliot MJ, et al. The role of airway stents in the management of pediatric tracheal, carinal, and bronchial disease. *Eur J Cardiothorac Surg*. 2000;18:505-12.
4. Furman RH, Backer CL, Dunham ME, Donaldson J, Mavroudis C, Holinger LD. The use of balloon-expandable metallic stents in the treatment of pediatric tracheomalacia and bronchomalacia. *Arch Otolaryngol Head Neck Surg*. 1999;125:203-7.
5. Matt BH, Myer CM III, Harrison CJ, Reising SF, Cotton RT. Tracheal granulation tissue. A study of bacteriology. *Arch Otolaryngol Head Neck Surg*. 1991;117:538-41.
6. Santoro G, Picardo S, Testa G, Formigari R, Marianeschi S, Catena G, et al. Balloon-expandable metallic stents in the management of tracheomalacia in neonates. *J Thorac Cardiovasc Surg*. 1995;110:1145-8.
7. Nicolai T. Airway stents in children. *Pediatr Pulmonol*. 2008;43:330-44.
8. Zajac A, Krysta M, Kiszka A, Gorecki W. Biodegradable airway stents: novel treatment of airway obstruction in children. *Adv Clin Exp Med*. 2019;28:961-5.
9. Roscoe A, Kanellakos GW, McRae K, Slinger P. Pressures exerted by endobronchial devices. *Anesth Analg*. 2007;104:655-8.
10. Freitag L, Darwiche K. Endoscopic treatment of tracheal stenosis. *Thorac Surg Clin*. 2014;24:27-40.
11. Mondal A, Ha J, Jo VY, Wu F, Kaza AK, Dupont PE. Preclinical evaluation of a pediatric airway stent for tracheobronchomalacia. *J Thorac Cardiovasc Surg*. 2021;161:e51-60.